



Genvinset[®]

HFE H63D

INSTRUCTIONS FOR USE

Kit for detection of the H63D mutation of the HFE gene

For In Vitro Diagnostic use

Rev. 06 / 2022-07-07



Camino del Pílon 86, Casa 7, Local
50011 – Zaragoza (Spain)



www.bdrdiagnostics.com



Product codes:

GVS-H63D-24 (24 tests)

GVS-H63D-48 (48 tests)

GVS-H63D-96 (96 tests)

UDI-DI:

8437016942178

8437016942185

8437016942734

Store

from -30°C to -18°C

Genvinset[®]

HFE H63D

Index

1- Information for safety	3
2- Intended Use	3
3- Summary and explanation	3
4- Procedure principles	4
5- Kit contents	5
6- Kit storage	5
7- Materials required but not supplied	5
8- Sample collection and preparation	6
9- Usage procedures	6
10- Results	7
11- Quality control	9
12- Specific operation data	10
13- Procedure limitations	12
14- Troubleshooting guide	12
15- References	14
16- Notice to purchaser	14
17- Changes control	14
18- Explanation of symbols used on the labels	15

1- Information for safety

Please, read completely these instructions for use and follow them during the use of the current IVD kit.

The IVD kit shall be used by experts with strong experience in DNA analysis and interpretation of genetic results.

Consult the manufacturer if there are doubts involving the description of the assay method. Contact by phone +34 976 094 603 or email address customersupport@bdrdiagnostics.com.

The IVD kit has a limited shelf life. Make sure that the shelf life is not expired before using the kit. Reagents from a kit beyond its expiry date might be degraded, which could impair results. Discard the expired reagents following applicable regulations.

The contamination of the sample or the reagent can render incorrect results. Be careful during the DNA extraction and sample and reagents manipulation.

This kit can be damaged during transport or storage. Do not use the kit in case of suspected damage during transport. Follow carefully the storage conditions described in the label and the IFU.

Ensure that the waste is managed according to local regulations. Incorrect waste management can result in environmental contamination.

The toxicological properties of the kit have not been studied in-depth, so it is recommended to avoid contact with skin and mucous membranes. Do not ingest. Manual Safety Data Sheets (MSDS) are available for the customers by request.

Ensure that this kit is adequate for the requested analysis by the clinician.

2- Intended Use

Genvinset® HFE H63D is a semi-automated *in vitro* diagnostic kit for the qualitative detection of the H63D mutation (NCBI dbSNP rs1799945; NM_000410.4: c.187C>G), in the *HFE* gene (OMIM: 613609) associated with hereditary hemochromatosis, in genomic DNA extracted from whole blood using Real Time PCR technology with specific TaqMan® probes.

Patients who can benefit from this determination are those referred by a specialist. The results of this test should not be the only ones on which the therapeutic decision is based and should be used as an aid in the diagnosis together with results of other markers of the disease.

The intended user of the kit is technical personnel trained to carry out the protocol and the interpretation of results described in this document.

3- Summary and explanation

Hereditary hemochromatosis (HH) is an inherited autosomal recessive disorder of iron metabolism. Due to excessive intestinal absorption, iron accumulates in the parenchymal cells of the liver, pancreas, heart, and other organs resulting in structural injury and impaired function¹. HH is one of the most common genetic diseases in Caucasians with a prevalence of 1 in 300 to 500². The symptoms are often non-specific, and the organ damage is often irreversible once it has occurred. Therefore, early detection and treatment are critical as part of preventive medicine³.

The discovery of the *HFE* gene in 1996 resulted in the inclusion of its molecular analysis in the diagnostic strategy for HH. The *HFE* gene is located on the short arm of chromosome 6, and it encodes the HFE protein, a 343 amino acid glycoprotein similar to MHC class I-type proteins. Three mutations in the *HFE* gene (C282Y, H63D, and S65C) have been associated with the development of iron overload, leading to a clinical diagnosis of HH^{1,4,5}.

Most HH cases (52-96%) in the European regions are associated with a homozygous mutation at position 845 (G→A) in exon 4 of the *HFE* gene (rs1800562), which results in an amino acid change at position 282 from cysteine to tyrosine (C282Y). A second mutation that is frequently detected occurs at position 187 (C→G) in exon 2 of the *HFE* gene (rs1799945), wherein the amino acid histidine is replaced by aspartic acid at position 63 (H63D). The contribution of this allele to iron overload is most relevant in the case of combined heterozygosity with allele C282Y (C282Y/H63D). The third *HFE* mutation is a substitution at position 193 (A→T) of exon 2 (rs1800730) resulting in an amino acid change at position 65 from serine to cysteine (S65C)^{1,4,5}. This mutation has proven to be generally benign, although the C282Y/S65C genotype can increase the risk of disease, contributing to a mild phenotype⁶.

4- Procedure principles

The test is based on real-time PCR technology with TaqMan[®] probes. Each sample is analysed with:

- A pair of primers for the amplification of a fragment of the *HFE* gene where the H63D mutation is located.
- A hydrolysis probe specific for the wild-type allele (C at rs1799945) labelled at the 5' end with HEX fluorophore and a hydrolysis probe specific for the mutant allele (G at rs1799945) labelled at the 5' end with FAM fluorophore. Both probes are labelled at the 3' end with a quencher that suppresses the fluorescence of the fluorophores when the probe is intact.

As the PCR reaction proceeds, the 5'→3' exonuclease activity of Taq polymerase cleaves the probes attached to their complementary sequence, separating fluorophore from quencher, and producing a fluorescent signal (in real-time), which is proportional to the amount of PCR product generated and monitored in a real-time PCR instrument. Thus:

- In the case of homozygous wild-type (wt) samples for the analysed change (C/C at rs1799945), the HEX-labelled wt allele-specific probe binds to its complementary sequence of the amplified gene, and the following is observed:
 - fluorescent signal in the HEX channel (λ_{\max} 556 nm), and
 - no signal or weak signal in the FAM channel (λ_{\max} 518 nm).
- In the case of homozygous mutant (mut) samples for the analysed change (G/G at rs1799945), the FAM-labelled mut allele-specific probe binds to its amplified complementary sequence. In this case, the following is detected:
 - fluorescent signal in the FAM channel and
 - no signal or weak signal in the HEX channel.
- In the case of heterozygous samples for the analysed change (C/G at rs1799945), both probes shall bind to the amplified DNA sequences generating:
 - FAM channel signal and
 - signal in the HEX channel.

5- Kit contents

→ GVS-H63D-24 (24 tests)

- GVS-H63D-PM: 1 vial x 96 µL Primer Mix (PM) - Blue cap
- GVS-HFE-MM: 1 vial x 120 µL Master Mix (MM) - Red cap
- GVS-H63D-C1: 1 vial x 15 µL Control WT (C1) - Green cap
- GVS-H63D-C2: 1 vial x 15 µL Control MUT (C2) - Green cap with an orange sticker
- GVS-RB: 1 vial x 100 µL Reaction Blank (RB) - Natural cap

→ GVS-H63D-48 (48 tests)

- GVS-H63D-PM: 2 vials x 96 µL Primer Mix (PM) - Blue cap
- GVS-HFE-MM: 2 vials x 120 µL Master Mix (MM) - Red cap
- GVS-H63D-C1: 1 vial x 15 µL Control WT (C1) - Green cap
- GVS-H63D-C2: 1 vial x 15 µL Control MUT (C2) - Green cap with an orange sticker
- GVS-RB: 1 vial x 100 µL Reaction Blank (RB) - Natural cap

→ GVS-H63D-96 (96 tests)

- GVS-H63D-PML: 1 vial x 460 µL Primer Mix (PML) - Blue cap
- GVS-HFE-MML: 1 vial x 575 µL Master Mix (MML) - Red cap
- GVS-H63D-C1L: 1 vial x 50 µL Control WT (C1L) - Green cap
- GVS-H63D-C2L: 1 vial x 50 µL Control MUT (C2L) - Green cap with an orange sticker
- GVS-RB: 1 vial x 100 µL Reaction Blank (RB) - Natural cap

6- Kit storage

All the components of the kit must be stored between -30°C and -18°C upon receipt. Under these conditions, they are stable up to their expiry date.

Do not perform more than 3 freeze/thaw cycles on the kit vials as this may reduce the sensitivity of the assay and affect the results. If assays are to be performed with few samples, it is recommended to use aliquots of the reagents to reduce freeze/thaw cycles.

Due to the photosensitive nature of the Primer Mix, avoid continuous exposure to light.

7- Materials required but not supplied

General

- Disposable gloves
- Lab coat

Consumables

- Filter tips (P200, P20 and P10)
- 1.5 mL autoclaved tubes
- Compatible consumables for the real-time PCR instrument

Equipment

- Vortex mixer

- Centrifuge
- Micropipettes (P200, P20 and P10)
- Real-time PCR instrument, with FAM and HEX/VIC detection channels. The following devices have been validated:
 - StepOne™, 7500, QuantStudio™ 5 Dx, and QuantStudio™ 6 Flex Real-Time PCR systems, Applied Biosystems™
 - LightCycler® 96 System, and LightCycler® 480, Roche
 - DT Lite Real-Time PCR System, DNA-Technology
 - qTOWER³ G, Analytik Jena

8- Sample collection and preparation

Samples must be collected in line with the instructions for use of the collection device (not included with the kit) and under national and international guidelines.

The present test should only be performed with genomic DNA extracted from whole blood samples preserved with anticoagulant agents such as EDTA or citrate. Heparin can interfere with the PCR process and should be avoided.

The technique is compatible with several DNA extraction methods. Before delivering results with a diagnostic purpose, a validation assay with such extraction method should be done.



CAUTION!

All biological and blood samples should be treated as potentially infectious. When manipulating them, take the corresponding basic and universal precautions.

9- Usage procedures

→ PCR setup



CAUTION!

- Define pre- and post-PCR work areas that should be kept separated to reduce the risk of contamination. Prepare the PCR in the pre-PCR area. Use lab coat and disposable gloves.
 - Work on ice or over a cool block. Minimize the time between preparation and start of the assay.
 - For each assay, it is recommended to test the contamination control (Reaction Blank) and both the Control WT (C1) and Control MUT (C2) included in the kit.
1. Thaw all the kit components before starting the assay. Vortex vigorously the Primer Mix vials and mix carefully the Master Mix vials. Centrifuge briefly to collect the volume at the bottom of the tubes.
 2. Prepare the reaction mix for n+1 samples, using the quantities indicated in the following table.

	Vol. per sample (µL)
Master Mix	5
Primer Mix	4

- Gently mix and centrifuge to ensure that all volume settles to the bottom of the tubes.
- Pipette 9 µL of this mix into the PCR plate/tubes.
 - Add into each well 1 µL of DNA, Reaction Blank, Control WT or Control MUT.
 - Seal the plate/tubes using the appropriate sealer and centrifuge briefly to remove any bubbles. Ensure that all the volume settles to the bottom of the well.
 - Place the plate/tubes in the thermal cycler and set up the thermal cycler amplification program as described in the following section.

→ Thermal cycler configuration

- Set up the following readout channels:
 - FAM channel for detection of the mut probe.
 - HEX/VIC channel for detection of the wt probe.
- Set up the following amplification program and start the run:

	Cycles	Temperature (°C)	Time (mm:ss)	Analysis
Denaturation	1	95	05:00	X
Cycles	50	95	00:15	X
		64	01:00	Single
Cooling	1	15	∞	X

→ Disposal

Waste products shall be managed according to local regulations.

10- Results

→ Results visualization

The analysis of the results is performed with the specific software of the real-time PCR instrument being used, and according to the manufacturer's instructions for use.

Once the run is finished, select the linear scale to visualize the amplification curves. It is recommended to check the correct behaviour of the obtained amplification curves:

- An amplification signal is considered positive if a quick and regular increase of fluorescence values (exponential) is observed (sigmoidal amplification) with Ct<35.
- A weak fluorescent, background signal or exponential signal with Ct>35 should not be considered a positive amplification. This assay allows the detection of alleles that differ by only one nucleotide and, therefore, in samples homozygous for one of the alleles, weak non-specific signals from the fluorophore used for the detection of the other allele can be observed. The appearance of these signals does not invalidate the assay.

A sample is considered positive if it produces an exponential amplification with $Ct < 35$. A sample is considered negative if it produces a non-exponential amplification with low intensity or an exponential amplification with Ct value > 35 .



CAUTION!

To determine the Ct value in each channel, adjust the threshold line as follows:

Select linear scale view and choose the region where the fluorescence signal is stable before exponential amplification. Place the threshold line above this background signal, so that it crosses close to the inflection point of the amplification curve. This line should slightly exceed the value of highest fluorescence obtained with negative samples for the allele detected in this channel.


→ Interpretation of results

The results obtained with this kit can be interpreted by visualization of the amplification curves in the FAM and HEX channels or by scatter plot of the end-point intensity values of the FAM and HEX channels.

Amplification curve


Select the "linear scale" and check the absence/presence of sigmoidal amplification curves in each channel.

Homozygous wt sample (C/C at rs1799945)




No signal or low intensity signal in FAM channel and exponential amplification in HEX channel ($Ct < 35$)

Homozygous mut sample (G/G at rs1799945)



Exponential amplification in FAM channel ($Ct < 35$) and no signal or low intensity signal in HEX channel


Heterozygous sample (C/G at rs1799945)



Exponential amplification signal ($Ct < 35$) in both FAM and HEX channels

Scatter plot

Many real-time PCR software programs allow to automatically plot the fluorescence intensity data of one channel versus the other (allelic discrimination/genotyping). In this type of representation, data points located close to the X and Y axes correspond to homozygous genotypes for the allele detected with the fluorophore represented on the corresponding axis. Points located approximately in the middle of the axis correspond to heterozygous genotypes and the negative control (Reaction Blank) should appear at the bottom left, close to the origin.




Plot showing heterozygous (green dots), homozygous mut (blue dots), and homozygous wt samples (red dots) for the H63D mutation and a Reaction Blank (black cross), using Genvinset[®] HFE H63D kit.

11- Quality control

The kit includes a Reaction Blank, a Control WT (C1) and a Control MUT (C2) which must be tested in each assay. An adequate behaviour of these control samples is a guarantee of the proper performance of the reaction.


The results are considered valid if the following amplification pattern is observed in the control samples:

Control WT (C1)




No signal or low intensity signal in FAM channel and sigmoidal signal in HEX channel ($Ct < 35$)

Control MUT (C2)



Amplification signal in FAM channel ($Ct < 35$) and no signal or low intensity signal in HEX channel

Reaction Blank



No signal in both FAM and HEX channels or amplification with $Ct > 35$

The results are considered invalid and should be repeated if:

- An amplification curve with $Ct < 35$ is observed in FAM and/or HEX channels in the Reaction Blank.
- Non-exponential amplification signal is observed or an amplification signal with $Ct > 35$ appears in the Control WT or Control MUT wells.

If adequate behaviour is observed in the positive control reactions, proceed with the interpretation of the rest of the samples as indicated in the previous section.

12- Specific operation data

➔ Analytical specificity

Cross-reactivity has been measured in three independent Genvinset® HFE H63D kit validation studies that are described below.

In addition, primers and probes alignment have been checked *in silico*. The alignment of primer sequences is specific. Probes align specifically at the rs1799945 position (NM_000410.4: c.187C>G) at the *HFE* gene (OMIM: 613609). No cross-reaction phenomena with genomic DNA have been reported.

Interference has been studied by means of a bibliographic search. Heparin can inhibit Taq polymerase activity and compete with target nucleic acid, so the blood collected must be treated with other anticoagulants as stated in the "Sample collection and preparation" section. Some substances present in blood are known as PCR inhibitors: haemoglobin, hemin, bilirubin, bile salts, lactoferrin, and immunoglobulin G. The polymerase included in Genvinset® HFE H63D has demonstrated high resistance to inhibition, and Master Mix composition is designed to not be affected by interference substances. Nevertheless, the presence of potentially inhibitory substances must be eliminated during DNA extraction and purification protocol. Before delivering results with a diagnostic purpose, a validation assay with such extraction method should be done.

→ Analytical sensitivity

An assay using 1:4 serial dilutions of two DNA samples, one homozygous for H63D and one heterozygous for H63D was performed. DNA samples were obtained by conventional extraction system at a concentration of 72.26 and 27.36 ng/μL respectively. The following results were obtained in terms of analytical sensitivity in the detection of wild-type and mutant alleles:

- Detection Limit of wild-type and mutant alleles = 0.03 ng/μL (Ct<35)

→ Diagnostic sensitivity and specificity

97 DNA samples obtained from two clinical laboratories were analysed. They were previously genotyped by different methodologies.

All the samples tested were considered valid (positive amplification of at least one allele) except for one in which both FAM and HEX amplifications failed. The following results were obtained:

		Genvinset® HFE H63D		
	Genotype	wt/wt (C/C at rs1799945)	wt/mut (C/G at rs1799945)	mut/mut (G/G at rs1799945)
Previous method	wt/wt	50	0	0
	wt/mut	0	33	0
	mut/mut	0	0	13

There is a 100% match in the results obtained with Genvinset® HFE H63D and the genotyping previously obtained with other different methods used in each laboratory.

→ Accuracy

Repeatability study consists in the assessment of within-run variability, through the analysis of replicas of every kind of sample that can be measured by the kit (homozygous and heterozygous samples). Each sample was analysed in duplicate.

Genvinset® HFE H63D showed a 100% repeatability.

A study to determine the reproducibility of the reagent was performed. It allows to estimate the variability between runs, lots and operators. Samples used represent the full range of expected analyte that can be measured with Genvinset[®] HFE H63D, that is, homozygous wild-type (CC), homozygous mutant (GG) and heterozygous (CG) at rs1799945. Three operators ran the assay in three different runs and tested three different reagent lots.

Genvinset[®] HFE H63D showed a 100% reproducibility.

→ **Trueness**

The trueness of the analytical procedure of Genvinset[®] HFE H63D is assessed by comparison with a reference method. The study was developed as an internal validation of the reagent, in which trueness has been demonstrated with a 100% value. See the "Diagnostic Sensitivity and Specificity" section.

13- Procedure limitations

- The method detects the presence of C/G at rs1799945.
- Mutations or polymorphisms at annealing primer/probe sites are possible and may result in the lack of allele definition. Other technologies could be necessary to resolve the typing.
- All the recommendations mentioned in this document should be carefully followed. Any performances that do not meet such indications, can lead to poor results.
- Do not use the kit if there are any suspicions of possible loss of reactivity, contamination, external box deterioration or any other incidence that might affect the kit's performance.
- All Genvinset[®] reagent manipulations must be done according to good laboratory practice, being adapted to local regulations.
- Do not mix components from other kits or lot numbers.
- Do not use the kit after its expiry date. Dispose of expired reagents according to applicable regulations.
- The real-time PCR thermal cycler must be calibrated according to the manufacturer's recommendations and should be used accordingly to the manufacturer's instructions.
- Data and result interpretation should be revised by qualified personnel.
- This product is an auxiliary tool for the diagnosis of patients with suspected hereditary hemochromatosis. Use these results in conjunction with clinical data and results of other tests performed on the patient.

14- Troubleshooting guide

→ **No amplification signal is detected in any sample (neither in positive controls) or the intensity is very low**

- The PCR instrument is not correctly programmed. The thermal profile is not correct/the reading channels are not correctly configured/the selected fluorophores are not suitable.
 - Check that the instrument has been programmed correctly.
- The positions of the samples and controls indicated on the worksheet do not coincide with the positions in which they have been placed in the device.
 - Correctly assign the position of the samples.
- The reagent does not work properly.

- Ensure that the kit is stored at an appropriate temperature (between -30°C and -18°C) and protected from light. Avoid unnecessary freeze/thaw cycles. Do not use it beyond the expiry date.
- The indicated amounts of each of the reagents have not been added to the reaction mix.
 - Check the volumes of each component added to the mix.
- The consumables used are not compatible with the equipment being used.
 - Make sure that you are using the correct consumables (compatible with the PCR instrument used).

→ No signal is detected in clinical samples (signal appears in positive controls)

- Poor quality of the DNA used.
 - Check the Absorbance ratio 260/280 and discard poor-quality samples. Avoid the presence of inhibitors (heparin, haemoglobin, hemin, bilirubin, bile salts, lactoferrin, immunoglobulin G). Repeat sampling and DNA extraction.
- Inadequate DNA concentration.
 - Adjust the DNA concentration to the recommended concentration range.
 - Check that the DNA concentration is above the limit of detection.
- Inhibition of amplification.
 - Collect whole blood in EDTA or citrate tubes.
- No sample was added.
 - Repeat assay making sure that samples have been added.

→ Signal detected in the negative control

- Dispensing error (pipetting).
 - Change the pipette tip each time DNA is added to a well. Check that the sample added to the well corresponds to what is written on the worksheet.
- Contamination of the Primer Mix/Master Mix/Reaction Blank vial.
 - Repeat the assay with fresh aliquots.
- The PCR preparation area is contaminated.
 - Clean surfaces, instruments, lab coats, and change consumables and reagents. Repeat the assay.

→ Fluorescence intensity varies between samples or abnormal amplification curves are observed

- Dirtiness outside the reaction tube interferes with the detection of fluorescence.
 - Clean the equipment thoroughly. Check that the outside of the tubes/plate is clean. Handle the plates/tubes with gloves.
- The volume is not at the bottom of the well or there are bubbles.
 - Centrifuge the plates/tubes before inserting them into the thermal cycler.
 - Check if there are bubbles. If so, remove them.
- The plate/tubes have not been closed properly.
 - Repeat the assay checking that the plates/tubes have been sealed correctly.
- DNAs with different concentrations have been used or the sample used contains a reaction inhibitor.
 - Repeat the sampling and DNA extraction.
- Presence of polymorphisms or mutations at the probe/primer binding sites.
 - Contact our Technical Support department through customersupport@bdrdiagnostics.com.

15- References

- 1) Katsarou, M. S., Papisavva, M., Latsi, R. & Drakoulis, N. Hemochromatosis: Hereditary hemochromatosis and HFE gene. *Vitamins and Hormones* **110**, 201–222 (2019).
- 2) Porter, J. L. & Rawla, P. Hemochromatosis. *Medicina Interna de Mexico* **35**, 896–905 (2021).
- 3) Pearlman, B. L. Hereditary hemochromatosis: Early detection of a common yet elusive disease. *Consultant-greenwich-*, **42**(2), 237–241 (2002).
- 4) Cukjati, M., Vaupotič, T., Ruprecht, R. & Čurin-Šerbec, V. Prevalence of H63D, S65C and C282Y hereditary hemochromatosis gene mutations in Slovenian population by an improved high-throughput genotyping assay. *BMC Medical Genetics* **8**, 69 (2007).
- 5) Kowdley, K. v., Brown, K. E., Ahn, J. & Sundaram, V. ACG Clinical Guideline: Hereditary Hemochromatosis. *American Journal of Gastroenterology* **114**, 1202–1218 (2019).
- 6) Holmström, P., Marmur, J., Eggertsen, G., Gåfvæls, M. & Stål, P. Mild iron overload in patients carrying the HFE S65C gene mutation: a retrospective study in patients with suspected iron overload and healthy controls. *Gut* **51**, 723 (2002).

16- Notice to purchaser












- This product has been developed for *in vitro* diagnostic purposes.
- When the kit is used within the territory of any Member State of the European Union, should there be any serious incident regarding the use of the kit, the user must report it to the manufacturer (regulatory@bdrdiagnostics.com) and the Competent Authority of their country and/or the patient's country. Serious incident reporting for all other countries must be conducted according to local requirements for each country.
- Blackhills Diagnostic Resources, S.L.U. products should not be resold, modified for reselling or used to manufacture other commercial products without the written consent of Blackhills Diagnostic Resources, S.L.U.
- All information contained in this document can suffer modifications without prior notice. Blackhills Diagnostic Resources, S.L.U. does not assume any responsibility for possible errors in the document. This document is considered complete and accurate at the time of its publication. In no case will Blackhills Diagnostic Resources, S.L.U. be responsible for accidental, special, multiple, or derived damages from the use of this document.
- The purchase of this product grants rights to the purchaser under certain Roche patents only used to provide *in vitro* diagnostic services. It does not grant any generic patent, or any other patents aimed at any other usage apart from the one specified.
- FAM™ and HEX™ are trademarks of Life Technologies Corporation.
- FAM™ and HEX™ may be covered by one or more patents owned by Applied Biosystems, LLC. The purchase price of this product includes limited, nontransferable rights.
- TaqMan® is a registered trademark of Roche Molecular Systems, Inc.
- Genvinset® is a trademark of Blackhills Diagnostic Resources, S.L.U.

17- Changes control

Version	Description of the modification
Rev. 05	Change of control C1 and C2 volume to 15µL.

Rev. 06	Correction of typos and translation mistakes. New sections: Accuracy and Trueness. Information regarding intended user, intended patient and interferences has been added. Change in Control WT and Control MUT name. Change in the list of real-time PCR instruments. Insertion of UDI-DI codes.
---------	---

18- Explanation of symbols used on the labels

	<i>In vitro</i> diagnostic medical device		Expiration date
	Catalogue number		Contents sufficient for <n> tests
	Lot number		Manufacturer
	Temperature limit		Keep away from sunlight
	Positive control		Consult electronic Instructions For Use document
	This product fulfils the requirements of Directive 98/79/EC on <i>in vitro</i> diagnostic medical device		